

SARTORIUS

Date and time	01/07/2025	17:36:24		
Calibration function	Cal.fuc.	Internal calibration/adjustment		
Start calibration	Start	Manual start		
Calibration deviation	Cal.dev.		+0.01	g
Adjustment function	Adj.fun.	Internal adjustment		
Adjustment deviation	Adj.dev.		+0.00	g