



Manufacturer	Manuf.	Sartorius
Model number	Model	MCA6202S-2S00-0
Serial number	Ser.no.	0037302471
Package version	Package	09-03-09.01.07 CN:1CBD
QAPP version	QAPP	2.0.1
Task name	Task	Test 2 Leopharma
Device ID 1	ID 1	LEOPHARMA
Device ID 2	ID 2	PACKAGING DEPT
Device ID 3	ID 3	LINEA XX

Initial:

Product ID	Prod. ID	LOCOBASE-REPAIR 100 g
Batch No.:	LOTTO:	final-test
Start	Start	10/06/2025 18:53:22
End	End	10/06/2025 18:55:36

Result:

Samples total	Total	15
Samples below tolerance 2	<TU2	0
Samples below tolerance 1	<TU1	0
Samples below tolerance	<TU	0
Samples OK	OK	15
Samples above tolerance	>TO	0
Samples above tolerance 1	>TO1	0
Samples above tolerance 2	>TO2	0

Statistics:

Number of components	n	15	
Mean value	$\bar{x}$	99.993	g
Long term mean value	LMean	99.994	g
Standard deviation	s	0.006	g
Variation coefficient	sRel	0.01	%
Sum of all values	Sum	1499.90	g
Lowest component value (minimum)	Min	99.98	g
Highest component value (maximum)	Max	100.00	g
Difference min. and max.	Diff	0.02	g

Date/time of report end	20/06/2025	16:50:17
Signat.:		